

TEST AND CALIBRATION CERTIFICATE

DATE CERTIFICATE N^o.

LOAD CELL TYPE SERIAL N^o.

CALIBRATION METHOD CALIBRATION MODE

BRIDGE RESISTANCE INSULATION RESISTANCE

(NOMINAL)

CONNECTIONS

+VE SUPPLY	RED	-VE SUPPLY	BLUE
+VE OUTPUT	GREEN	-VE OUTPUT	YELLOW
	SCREEN	BLACK	

PROOF LOAD

LOAD TONNE	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
17.00	0.5630	0.5476	0.56
34.00	1.1155	1.0951	0.74
51.00	1.6565	1.6427	0.50
68.00	2.1875	2.1903	-0.10
85.00	2.7175	2.7379	-0.74

ZERO OFFSET mV/V CAL. READING mV/V CAL. RESISTOR VALUE KΩ CAL. EQUIVALENT CALIBRATED BY:WEBMASTER**ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS***Registered Office and Trading Address:*

LCM Systems Limited • Unit 15, Newport Business Park • Barry Way • Newport • Isle of Wight • PO30 5GY • UK
 Tel: +44 (0)1983 249264 • Fax: +44 (0)1983 249266 • Email: info@lcm systems.com

Http: www.lcm systems.com